

PF-3 COD (RM0934302)

NANOCOLOR Test	Intervallo di misura	Numero Test	Numero Determinazioni	Codice N°
COD 40	2–40 mg/L O ₂	0–27	20	CM0985027
COD 60	5–60 mg/L O ₂	0–22	20	CM0985022
COD 160 ISO 15705	15–160 mg/L O ₂	0–26	20	CM0985026
COD 600 ISO 15705	50–600 mg/L O ₂	0–30	20	CM0985030
COD 1500 ISO 15705	100–1500 mg/L O ₂	0–29	20	CM0985029
COD 4000	400–4000 mg/L O ₂	0–11	20	CM0985011
COD 10000	1.00–10.00 mg/L O ₂	0–23	20	CM0985023
COD 15000	1.0–15.0 mg/L O ₂	0–28	20	CM0985028
COD 60000	5.0–60.0 mg/L O ₂	0–12	20	CM0985012